

**National Textiles & Garment Quality Supervision
Inspection Centre (Zhejiang)
Test Report**

ADDRESS: No.300 XIASHA ROAD
HANGZHOU CHINA

Tel: (0571) 85122669

Fax: (0571) 85125121

Postcode: 310018

NUMBER:W202004379E


DATE: 2020-04-08

APPLICANT: Zhejiang Hongao Medical Articles Co.,Ltd

SAMPLE DESCRIPTION: Civilian use KN95

ADDRESS: No. 951. Xinghe Road, Wanquan Town, Pingyang County, Wenzhou City, Zhejiang Province



ITEM	STANDARD (KN95)	RESULT	RATING
1. FILTER EFFICIENCY(SALT MEDIUM) (GB 2626-2006)(%)			
-	≥95.0	96.3	PASS
2. RESPIRATORY RESISTANCE (GB 2626-2006)(Pa)			
-INSPIRATORY RESISTANCE	≤350	166	PASS
-EXPIRATORY RESISTANCE	≤250	123	PASS
PICTURE			

Remarks:The sample is not pretreated.Result for reference.

俞杰

PREPARED AND CHECKED BY

YU JIE

----END----